ORIGINAL ARTICLE

Digital evaluation of the reproducibility of implant scanbody fit—an in vitro study

Michael Stimmelmayr • Jan-Frederik Güth • Kurt Erdelt • Daniel Edelhoff • Florian Beuer

Received: 17 November 2010 / Accepted: 10 May 2011 / Published online: 4 June 2011 © Springer-Verlag 2011

Abstract Dental restorations are increasingly manufactured by CAD/CAM systems. Currently, there are two alternatives for digitizing dental implants: direct intraoral data capturing or indirect from a master cast, both with transfer caps (scanbodies). The aim of this study was the evaluation of the fit of the scanbodies and their ability of reposition. At the site of the first molars and canines, implants were placed bilaterally in a polymer lower arch model (original model), and an impression was taken for fabricating a stone cast (stone model). Ten white-light scans were obtained from the original and the stone model with the scanbodies in place. The scanbodies were retrieved after each scan and re-attached to the same implant or lab analogue. The first scan of the series served as control in both groups. The subsequent nine scans and control were superimposed using inspection software to identify the discrepancies of the four scanbodies in both experimental groups. The systematic error of digitizing the models was 13 µm for the polymer and 5 µm for the stone model. The mean discrepancy of the scanbodies was 39 μ m (±58 μ m) on the original implants versus 11 μ m (±17 μ m) on the lab analogues. The difference in scanbody discrepancy between original implants and lab analogues was statistically significant (p < 0.05, Mann-Whitney U test). Scanbody discrepancy was higher on

M. Stimmelmayr (🖂)

Private Practice for Oral Surgery, Josef-Heilingbrunnerstrasse 2, 93413 Cham, Germany e-mail: michael.stimmelmayr@med.uni-muenchen e-mail: praxis@m-stimmelmayr.de

J.-F. Güth · K. Erdelt · D. Edelhoff · F. Beuer Department of Prosthodontics, University of Munich, Goethestrasse 70, 80336 Munich, Germany original implants than on lab analogues. Fit and reproducibility of the scanbodies on original implants should be improved to achieve higher accuracy of implant-supported CAD/CAM fabricated restorations.

Keywords Implants · Scanbodies · CAD/CAM · Impression · Model scanning

Introduction

High-precision transfer of the clinical situation into dental laboratory is one of the crucial factors for highly accurate prosthetics on natural teeth as well as dental implants. This transfer includes the implant position as well as the inclination [1, 2]. The absolute passive fit of the prosthodontics-the declared goal of any rehabilitation-is, however, ruled out by various sources of error [3]. For implant-supported restorations, discrepancies are particularly detrimental because of the rigid osseointegration of the implants. Those discrepancies may lead to both mechanical and biological complications. Stress-induced porcelain chippings, screw loosenings and fractures of the screw, abutment, or even the implant, were reported [4-9]. Biological complications as mucositis, periimplantitis [10–12], and implant loss caused by inappropriate loading were discussed, too [13].

The precision of intra-oral impressions is one of the most important factors to achieve a perfect fit [14]. Impression techniques and impression materials affect the precision of intra-oral data transfer [15, 16]. Conventional impressions are associated with transfer problems caused by shrinkage, variable layer thickness or separation of the impression material from the tray, and warping of the impression [17]. Additional problems are caused by the

expansion of the dental stone during cast fabrication and dimension changes by casting of the framework and veneering in the ceramic furnace.

In order to minimize sources of error, CAD/CAM techniques (computer-aided design/computer-aided manufacturing) have been introduced. For CAD/CAM-assisted fabrication, digitization of the clinical situation means a prerequisite. Two alternatives of data capturing are available: (1) intra-oral scanning (direct data capturing) or (2) digitizing the casts made from conventional impressions by scanning in the dental lab (indirect data capturing) [18].

Dental implants, unlike natural abutments, with their inner configuration defy digital capturing. Therefore, implant manufacturers offer devices, so called scanbodies, for digitizing of implants. These are clip or screw retained on the implants intra-orally or on the lab analogues of the master cast during scanning. The fit of these scanbodies is decisive for a high-precision transfer of the implant position and inclination, which is important for the fabrication of prosthodontics.

This study evaluated the accuracy of these scanbodies both on implants and lab analogues and detected potential sources of error. The working hypothesis is that the reproducible fit of the scanbodies will be the same on original implants and lab analogues.

Material and methods

Two-piece implants (Screwline Promote ø4.3/13 mm; Camlog Biotechnologies, Wimsheim, Germany) were placed at the sites of the first molars and canines bilaterally in an edentulous mandibular model (B-3 NM J UK; Frasaco, Tettnang, Germany). Impression posts were screwed onto them and a polyether impression (Impregum; 3 M-ESPE, Seefeld, Germany) was made with an open-bite custom tray (Master Impression Tray, Water Pik, Ft. Collins, CO, USA). Impressions remained on the model for 8 min, counted from the start of mixing [19]. Four hours after taking the impression, the cast was fabricated (Rocky Mountain Sahara; Klasse IV Dental GmbH, Augsburg, Germany) [19].

Cover screws (Camlog) were seated on the original implants in the Frasaco model (original model), which was matted by applying contrast spray (Met-L-Check Cleaner; Met-L-Check, Santa Monica, CA, USA). The applied contrast spray was used because of its control and better measurability. In unpublished pilot studies, the authors determined that Met-L-Check Cleaner made uniform layers of approximately 10 μ m, while other tested sprays produced layers of up to 100 μ m.

The cover screws were removed and scanbodies (Camlog) were seated on the implants clockwise with a defined torque

of 5 Ncm (Torque Control; Camlog) (personal communication with Camlog Biotechnologies), (Fig. 1). In a pilot study, the cast model was scanned five times without removing the scanbodies to determine the systematic error. These five data sets were superimposed and evaluated (Comet INSPECT[®] plus 4.5; Steinbichler Optotechnik, Neubeuern, Germany). The inspection software calculated discrepancies of 6 μ m between the data sets. This value represented the systematic error.

The original model was placed in a white-light scanner (Everest Scan Pro; KaVo, Biberach, Germany) for scanning the entire lower arch. When the first scan was completed, the object support was taken out of the scanner. The scanbodies were detached from the implants one by one and reattached on the same implant clockwise at a defined torque of 5 Ncm without removing the original model from the object support. Once all four scanbodies were relocated again, the object support was placed in the scanner and another scan was obtained for data acquisition. This was repeated nine times and the STL data (Standard Triangulation Language) of ten scans of the original model with the same scanbodies on the same implants were available. Those scanbodies were designed for a maximum of 20 uses (personal communication with Camlog Biotechnologies).

The same procedure was followed for the stone cast (cast model). As the dental stone used for the cast was specifically developed for the scanning technique, opaquing with scan powder was not required. Like the original model, the stone cast was left on the object support, while the scanbodies were released and relocated. Again, a series of ten scans were obtained, and data were stored in STL. The surface of each scanbody was defined by about 40.250 triangles.

Evaluating the positions of the scanbodies and their reproducible fit the STL data were imported and processed with an inspection software (Comet INSPECT[®] plus 4.5)



Fig. 1 Original polymer model with screw-retained scanbodies

for data comparison. The first scan model (control) was used as reference (Fig. 2) and compared with the remaining nine digital models. The original model remained the same, while the scanbodies were released and re-attached. Therefore, the scanbodies of the control were blanked out for superimposition with the nine digital models (target models 1-9) and only the model parts without the scanbodies were superimposed by the software. For optimal superimposing of the control and the target models (best fit), a three-run iterative approach (search radius 1, 0.5, and 0.1 mm and 5° search angle) was used. Additionally, the software (Comet INSPECT® plus 4.5) calculated the mean discrepancy of the control model and the target model for each superimposition. This represented the systematic error of the superimposition. The mean error of each superimposition was imported into a statistics program (SPSS 17.0, SPSS Inc, Chicago, IL, USA). The data of original models and lab casts were compared at a level of 5% (ANOVA). When the best fit was found and the value of the discrepancy was calculated and listed, the scanbodies were put back in the digitized control model, and the discrepancies of the four scanbodies in the nine target models were calculated. For this purpose, a plane was constructed through the upper third of the scanbodies on the control model (Fig. 3). In a distance of 0.6 mm, 21 measurement points were determined along the outline of each scanbody (Fig. 4), and the discrepancies between these particular 21 measurement points of the control and target were measured in this plane. The mean of these discrepancies were defined as scanbody discrepancy. This procedure was also used for the cast model with the lab analogues.

This provided nine comparisons of the four implant positions at the sites of the lower molars and canines for both the original model with the original implants and the cast model with the lab analogues. Twenty-one measuring



Fig. 2 Digitized original model with scanbodies



Fig. 3 Measuring plane constructed through upper third of scanbodies and digitally measured scanbody discrepancy

points were defined for each scanbody, i.e., 84 measuring points for each model. Overall, the nine comparisons yielded 756 measuring points for both the original implants and the lab analogues (Fig. 5).

Data were imported into a statistics program (SPSS 17.0, SPSS Inc). For explorative data analysis, the Mann–Whitney U test and ANOVA test was used. The level of statistical significance was set at 5%.

Results

Superimposition of control and target models without scanbodies exhibited discrepancies of 13 μ m (±3 μ m) for the original model and 5 μ m (±2 μ m) for the lab casts, respectively. Table 1 shows the discrepancies between the particular superimpositions of the models. There were statistically significant discrepancies of the superimposition between original and cast models (*p*≤0.001; ANOVA test).

The calculated mean discrepancies of the scanbodies were 39 μ m (±58 μ m) for the original model and 11 μ m (±17 μ m) for the stone model. Therefore, the scanbody discrepancies of the original implants differed significantly



Fig. 4 Measuring plane of the scanbody with determined 21 measurement points





from that of the lab analogues ($p \le 0.001$; Mann–Whitney U test). The mean discrepancies of the scanbodies at the different implant sites (FDI 36, 33, 43, and 46) for the original implants and the lab analogues are shown in Fig. 6.

Discussion

Scanbodies on lab analogues exhibited higher reproducibility of fit compared to original implants, rejecting the working hypothesis. Comparing the systematic error of the pilot study (6 μ m), which was done only on the stone model, with the calculated systematic error of the inspection software (5 μ m for the stone model), the results were almost equal. Therefore, the calculated systematic error of the original model (13 μ m) appears correct. Del Corso reported of a systematic error between 14 and 21 μ m, simulating an intra-oral data capturing in an in vitro simulation [20]. Apparently, the scans with the white-light scanner, described above, were quite exact. KaVo states a systematic error for the Everest Scan Pro scanning a complete jaw of $8-20 \ \mu m$ (personal communication). Also Mehl described the systematic error of extra-oral optical measurement systems for scanning stone casts of 20 μm or less [21].

The statistically significant superimposition discrepancy of 8 μ m between the original model and the stone model without scanbodies suggests that the scans of the original model are less precise. This might be due to the scan powder which was applied on the original model. However, scan powder is required for scanning reflecting surfaces like polymer models or most intra-oral scanning systems for opaquing the intra-oral surfaces [22].

The applied contrast spray was used because of its uniform layers of approximately 10 μ m. Other tested sprays produced layers of up to 100 μ m. This may be a reason, why the calculated systematic error of the digitizing of the original polymer model was less than the data reported by Del Corso and Mehl [20, 21]. However, this contrast spray is not allowed for intra-oral use, and scanners not requiring scan powder for digitizing objects would be desirable.

Original model	Discrepancy in mirometer (µm)	Stone model	Discrepancy in mirometer (µm)	Difference in mirometer (µm)
Control-target 1	8.005	Control-target 1	4.362	
Control-target 2	10.358	Control-target 2	4.102	
Control-target 3	12.234	Control-target 3	4.020	
Control-target 4	13.443	Control-target 4	3.618	
Control-target 5	13.933	Control-target 5	3.402	
Control-target 6	14.642	Control-target 6	9.176	
Control-target 7	15.126	Control-target 7	7.711	
Control-target 8	15.392	Control-target 8	6.308	
Control-target 9	15.948	Control-target 9	5.136	
Mean	13.231	Mean	5.315	7.916
Standard deviation	2.616	Standard deviation	2.008	

 Table 1
 Discrepancies of superimposition of original and cast models





855

For superimposition of the control and the target models, the scanbodies and parts of the alveolar crest of the control were blanked out. The "sublingual" and retromolar area showed the highest data quality and seemed to be most suitable for precise superimposition. Some parts of the alveolar crest showed bad edges and were not included in the superimposing process. The flat "sublingual" area gave an exact field for the space coordinates X and Y, the retromolar area for the space coordinates Z. Hence, the superimposition algorithm can be calculated precisely by the inspection software, shown by the low systematic error.

Comparing the mean discrepancies of the scanbody misfit on the original model/implant (39 µm) and on the stone model/lab analogue (11 µm), there is a statistically significant difference. Minus the systematic error of the digitizing of the models, the misfit on the original implant would be 26 μ m (39–13 μ m) versus 6 μ m (11–5 μ m) for the lab analogue. The accuracy of the interfaces of the original implants and the lab analogues should be identical (personal communication Camlog Biotechnologies). However, there is a production tolerance of about 15 µm. If the inner diameter of the lab analogues was on the lower limit and the original implants on the upper limit of manufacturing tolerance, a discrepancy of about 25 µm could be possible.

Ma described tolerances between the implant and the impression copings, abutment replicas, and abutments from 22 to 100 μ m [23]. These discrepancies are not only machining tolerances but also a function of the design of the positional index [24]. The positional index depends between internal and external implant-abutment connections [25]. Semper described rotational discrepancies of the abutment having three different geometric patterns of dental implants from 3.7° to 1.4° [26]. Based on these data, this study was conducted with an implant that showed an internal connection with less rotational discrepancy of 1.4°.

If multiple implants are parallel-inserted, there will be no horizontal shift in the transfer; if the implants are angulated. the rotational misfit leads to a horizontal discrepancy. An angulation of 20° and a rotational freedom of 1.5° can result in a horizontal misfit up to 127 µm [27].

Also the value of the discrepancy is dependent to the vertical extension of the scanbodies and the constructed plane for the measurement. The higher the plane is constructed in relation to the implant shoulder, the higher the measured discrepancy relating to the discrepancy in the internal connection. The plane for the measurements was determined in the upper third of the scanbodies. This vertical dimension showed approximately the length of a natural tooth, for determination the discrepancy in the occlusal plane.

Currently, it seems to be more precise to do a conventional impression and white-light scanning of stone casts for an accurate fit of implant-supported prosthodontics according to the results of this study. This is also stated by Luthardt [28]. However, it has to be considered that indirect data capturing consists of more steps in the working sequence. This includes the possibility of more discrepancies in the individual process.

Conclusions

Within the limitations of this study, the results showed that

- 1. the systematic error by scanning the stone models was less in contrast to the polymer models; therefore, whitelight scanning of stone casts seems to produces more accurate data than original (polymer) in vitro models matted with scan powder.
- 2. the ability of repositioning of the scanbody is better on lab analogues than on original implants.
- 3. the companies have to improve the reproducible fit of the scanbodies in the original implants by reducing the production tolerance.

Conflict of interest The authors declare that they have no conflict of interest.

References

- Assuncao WG, Filho HG, Zaniquelli O (2004) Evaluation of transfer impressions for osseointegrated implants at various angulations. Implant Dent 13:358–366
- Carr AB (1991) Comparison of impression techniques for a fiveimplant mandibular model. Int J Oral Maxillofac Implants 6:448– 455
- Kan JY, Rungcharassaeng K, Bohsali K, Goodacre CJ, Lang BR (1999) Clinical methods for evaluating implant framework fit. J Prosthet Dent 81:7–13
- Balshi TJ (1996) An analysis and management of fractured implants: a clinical report. Int J Oral Maxillofac Implants 11:660– 666
- Burguete RL, Johns RB, King T, Patterson EA (1994) Tightening characteristics for screwed joints in osseointegrated dental implants. J Prosthet Dent 71:592–599
- Eckert SE, Meraw SJ, Cal E, Ow RK (2000) Analysis of incidence and associated factors with fractured implants: a retrospective study. Int J Oral Maxillofac Implants 15:662–667
- Jemt T, Rubenstein JE, Carlsson L, Lang BR (1996) Measuring fit at the implant prosthodontic interface. J Prosthet Dent 75:314–325
- Sahin S, Cehreli MC (2001) The significance of passive framework fit in implant prosthodontics: current status. Implant Dent 10:85–92
- Wee AG, Aquilino SA, Schneider RL (1999) Strategies to achieve fit in implant prosthodontics: a review of the literature. Int J Prosthodont 12:167–178
- Augthun M, Conrads G (1997) Microbial findings of deep periimplant bone defects. Int J Oral Maxillofac Implants 12:106–112
- Leonhardt A, Renvert S, Dahlen G (1999) Microbial findings at failing implants. Clin Oral Implants Res 10:339–345
- Lindhe J, Berglundh T, Ericsson I, Liljenberg B, Marinello C (1992) Experimental breakdown of peri-implant and periodontal tissues. A study in the beagle dog. Clin Oral Implants Res 3:9–16
- Binon PP, McHugh MJ (1996) The effect of eliminating implant/ abutment rotational misfit on screw joint stability. Int J Prosthodont 9:511–519

- Karl M, Winter W, Taylor TD, Heckmann SM (2004) In vitro study on passive fit in implant-supported 5-unit fixed partial dentures. Int J Oral Maxillofac Implants 19:30–37
- Al-Bakri IA, Hussey D, Al-Omari WM (2007) The dimensional accuracy of four impression techniques with the use of addition silicone impression materials. J Clin Dent 18:29–33
- Lee H, So JS, Hochstedler JL, Ercoli C (2008) The accuracy of implant impressions: a systematic review. J Prosthet Dent 100:285–291
- Christensen GJ (2008) Will digital impressions eliminate the current problems with conventional impressions? J Am Dent Assoc 139:761–763
- Beuer F, Schweiger J, Edelhoff D (2008) Digital dentistry: an overview of recent developments for CAD/CAM generated restorations. Br Dent J 204:505–511
- Luthardt RG, Walter MH, Quaas S, Koch R, Rudolph H (2010) Comparison of the three-dimensional correctness of impression techniques: a randomized controlled trial. Quintessence Int 41:845–853
- 20. Del Corso M, Aba G, Vazquez L, Dargaud J, Dohan Ehrenfest DM (2009) Optical three-dimensional scanning acquisition of the position of osseointegrated implants: an in vitro study to determine method accuracy and operational feasibility. Clin Implant Dent Relat Res 11:214–221
- Mehl A, Ender A, Mormann W, Attin T (2009) Accuracy testing of a new intraoral 3D camera. Int J Comput Dent 12:11–28
- 22. Masek R (2005) Margin isolation for optical impressions and adhesion. Int J Comput Dent 8:69–76
- Ma T, Nicholls JI, Rubenstein JE (1997) Tolerance measurements of various implant components. Int J Oral Maxillofac Implants 12:371–375
- Semper W, Kraft S, Kruger T, Nelson K (2009) Theoretical considerations: implant positional index design. J Dent Res 88:725–730
- Binon PP (2000) Implants and components: entering the new millennium. Int J Oral Maxillofac Implants 15:76–94
- Semper W, Kraft S, Kruger T, Nelson K (2009) Theoretical optimum of implant positional index design. J Dent Res 88:731–735
- Semper W, Kraft S, Mehrhof J, Nelson K (2010) Impact of abutment rotation and angulation on marginal fit: theoretical considerations. Int J Oral Maxillofac Implants 25:752–758
- Luthardt RG, Loos R, Quaas S (2005) Accuracy of intraoral data acquisition in comparison to the conventional impression. Int J Comput Dent 8:283–294

Copyright of Clinical Oral Investigations is the property of Springer Science & Business Media B.V. and its content may not be copied or emailed to multiple sites or posted to a listserv without the copyright holder's express written permission. However, users may print, download, or email articles for individual use.